



Certificate of Analysis

Product Name: **PACE Multiplex Master Mix**
Product N°.: 007-0013, 007-0014, 007-0015 & 007-0016
Lot N°.: 82711213


Analytical Test

Taq level	PASS
pH @ 25°C	PASS
dNTP conc.	PASS
Magnesium (Mg ²⁺) conc.	PASS
Bacterial DNA level	PASS
DNase level	PASS
ROX level	PASS


Performance Test

3CR Bioscience PACE Multiplex Master Mix is functionally tested for allelic discrimination using several quality control assays (2 assays shown below) in pairs, comparing results to two previous lots of master mix (1 lot shown below) run with the same assays, at the same time, on the same instrumentation.

Assay 1
Lot: 82711213



Assay 1
Lot: 61720271



Assay 2
Lot: 82711213



Assay 2
Lot: 61720271



Result **PASS**

Quality Approved :

Date : 29th Nov 2022